



# Optilab

ILX Lightwave SSB-9200 High Density Source Bank Testing		
Wavelength Accuracy: Within +/- 0.4 nm; Wavelength Adj. Range: +/- 0.85 nm		
Channel #	Wavelength (nm)	Power Detected @ 10 dBm Setting (dBm)
1	1530.33	9.80
2	1531.12	9.96
3	1531.90	9.99
4	1532.68	9.45
5	1536.61	9.79
6	1534.25	9.54
7	1535.04	10.08
8	1535.82	10.04
9	1538.19	10.13
10	1538.91	9.60
11	1539.77	9.38
12	1540.56	7.27
13	1544.53	9.69
14	1542.14	9.21
15	1542.94	9.28
16	1543.73	10.16
17	1546.92	10.13
18	1547.72	9.93
19	1548.51	9.57
20	1546.12	9.65
21	1550.12	9.57
22	1550.92	9.90
23	1551.72	10.46
24	1552.52	10.58
25	1554.94	9.96
26	1555.75	9.51
27	1556.55	9.48
28	1554.13	9.41
29	1558.98	9.70
30	1558.98	9.70
31	1559.79	9.01
32	1560.61	9.52

Optilab LLC  
5110 N. 44<sup>th</sup> Street, Suite 200L  
Phoenix AZ 85018  
Phone.1.888.553.3888  
Fax.602.343.1489



Optilab LLC  
 5110 N. 44<sup>th</sup> Street, Suite 200L  
 Phoenix AZ 85018  
 Phone.1.888.553.3888  
 Fax.602.343.1489